

|          |  |          |  |
|----------|--|----------|--|
| ANALYST: |  | VPDES NO |  |
|----------|--|----------|--|

Parameter: Temperature

Method: Thermometric

05/04

METHOD OF ANALYSIS:

|  |  |
|--|--|
|  | 18th EDITION OF STANDARD METHODS-2550 B                                      |
|  | EPA METHODS OF CHEMICAL ANALYSIS-170.1                                       |
|  | TECHNIQUES OF WATER-RESOURCES INVESTIGATIONS OF USGS, BOOK 1, CHAP. D1, 1975 |

|   | Y | N |
|---|---|---|
| 1) Is a good mercury filled or dial type centigrade thermometer or thermistor used? <b>NOTE:</b> Non-mercury filled thermometers may be used as long as accuracy and gradation requirements are met.) [SM-1; 170.1-2.1]   |   |   |
| 2) Are the thermometers markings etched on the capillary glass? [SM-1]  |   |   |
| 3) Does the thermometer have a scale adequate to meet permit monitoring requirements? [Permit]  |   |   |
| 4) Is the liquid in the thermometer continuous with no air spaces? [Permit]   |   |   |
| 5) Is the thermometer immersed until a steady reading is obtained? [Permit; 170.1-5.1]  |   |   |
| 6) Is reference thermometer's calibration/verification current? NOTE: NIST/NIST-traceable thermometers must be within expiration date on manufacturer's calibration certificate or re-certified yearly. [SM-1; 170.1-3.1] |   |   |

PROBLEMS: